

Certificate(s) of Calibration

Certificate Number
 230328-07850-9761002

An ISO 9001 registered company.

REV: SVL 12/05/2018

Certificate Record 1 of 1

Printed 3/28/2023

Instrument Identification

Company:

Received Date: **Mar, 28 / 2023**

Calibration Date: **Mar, 28 / 2023**

Address:

Issue Date: **Mar, 28 / 2023**

Model: **07850-HypotULTRA 4, AC/DC/IR Hipot**

Recommended Due: **Mar, 27 / 2024**

Temperature: **(23+/- 5) DEG C**

Serial: **9761002**

Customer Instrument ID: **9761002**

Humidity: **(35+/-15) %RH**

Test Procedure #: **802-5001**

Technician: **Peter Nehmzow**

Order #: 0723175

Calibration Location:

Ikonix

Remarks: **As received, the instrument was found to be non-operational, and needed repair. The non-operational condition was due to corrupted firmware and output issues due to faulty parts. After service repair, calibration, and after test, the instrument is within manufacturer's tolerance limits.**

The instrument listed above has been calibrated using the standards traceable to the International System of Units (SI). Ikonix certifies that this instrument meets or exceeds manufacturer's tolerance limits as stated in the applicable test procedure (unless otherwise noted). The results reported herein relate only to the item calibrated. There was no sampling used in this calibration.

Ikonix quality system is registered to ISO 9001:2015.

This certificate shall not be reproduced except in full, without the written approval of Ikonix.

Calibration Standards

Inst ID#	Model No	Model Description	Cal Date	Due Date	NIST Traceability
AR 602	TM320	TEMPERATURE & HUMIDITY DATA LOGGER	07/08/2022	07/08/2023	2188910 THUNDER SCI DUAL PR HUMID GEN DUE CAL 07/31/22 2210965 THERMOHYGROMETER TM320 DUE CAL 01/31/23 2191690 THUNDER SCI DUAL PR HUMID GEN DUE CAL 10/31/22
AR 647	CTGB4-2	ATS EVOLUTION LOAD REV B	04/22/2022	04/22/2023	AR668, AR701, AR669, AR670, 676,
AR 560	VD15-16.5-A-AR-T	HIGH VOLTAGE DIVIDER	06/16/2022	06/16/2023	ROSS ID 9188, VD240-6.2-CIC-K-DM DUE 01/31/23 ROSS ID 9303, UT5H DUE 8/31/22 ROSS ID 9353, 3458A DUE 12/17/22 ROSS ID 9390, 34461 DUE 11/14/22
AR 665	34465A	6.5 DIGIT MULTIMETER	12/13/2022	12/13/2023	EVL764209 ASSET A457 FLUKE 5700A/EP3 DUE 12/31/23 EVL804477 ASSET A1151 FLUKE 5522A DUE CAL 6/3/23 447509-0142-447519 ASSET A1078 HP 3325B FUN GEN DUE 9/5/23

Certified by:

PN-106

Peter Nehmzow

28105 N Keith Drive, Lake Forest, IL 60045 Phone: 1-847-367-4671
 info@ikonix.com www.ikonix.com

Inspected by:

Robert Nordeen