



Product Service

Attestation of Conformity

No. N8A 15 12 87615 006

Holder of Certificate: Slaughter Company, Inc.28105 N. Keith Drive
Lake Forest, IL 60045-4546
USA**Product:** Laboratory Equipment
(Ground Bond Tester)**Model(s):** SLT264**Parameters:**

Rated voltage:	100-120/200-240 Vac
Rated frequency:	50/60 Hz
Rated power:	Max. 850 VA
Rated output:	3.0-10.0A/0-600 ohm, 10.1-30.0A/0-200 ohm, 30.1-40.0A/0-150 ohm
Protection class:	I

Tested according to:
EN 61010-1:2010
EN 61010-2-030:2010
EN 61010-031:2002/A1:2008

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: 611101501301**Date,** 2015-12-15
(Bill Lin)

CE After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

Page 1 of 1